



Product Service


CERTIFICATE

No. Z2 096885 0005 Rev. 00

Holder of Certificate: Xylem Analytics Germany GmbH
Peringerstr. 10
85055 Ingolstadt
GERMANY

Production Facility(ies): 096885

Certification Mark:



Product: Application Software

Model(s):

Data logger systems:
Winlog.validation, Winlog.med

Parameters:

Data logger system to read and process data for routine control and validation applications in the field of medicine, pharmacy and food

Tested according to:

DIN ISO/IEC 25051:2014 (Functionality)
DIN EN ISO 9241-110:2008 (Usability)
FDA 21 CFR Part 11

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the testing and certification regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 028-713192683-000
Valid until: 2021-08-31

Date, 2020-08-27

(Siegfried Muckenschnabl)